

EC CERTIFICATE

Number: 2149850CE01

Full Quality Assurance System

Directive 93/42/EEC on Medical devices, Annex II excluding (4)
(Devices in Class IIa, IIb or III)

Manufacturer:

Applied Biomedical Systems B.V.
Oxfordlaan 55
6229 EV Maastricht
The Netherlands

For the product category(ies)

Cardiac AF analyzer

DEKRA grants the right to use the EC Notified Body Identification Number illustrated below to accompany the CE Marking of Conformity on the products concerned conforming to the required Technical Documentation and meeting the provisions of the EC-Directive which apply to them:

0344

Documents, that form the basis of this certificate:

Certification Notice 2149850CN, initially dated 3 October 2012
Addendum, initially dated 3 October 2012

DEKRA hereby declares that the above mentioned manufacturer fulfils the relevant provisions of 'Besluit Medische Hulpmiddelen', the Dutch transposition of the Council Directive 93/42/EEC of June 14, 1993 concerning Medical devices, including all subsequent amendments. The manufacturer has implemented a quality assurance system for design, manufacture and final inspection for the above mentioned product category in accordance to the provisions of Annex II of Council Directive 93/42/EEC of June 14, 1993 and is subject to periodical surveillance. For placing on the market of Class III devices an additional EC design examination certificate according to Annex II (4) is mandatory. The necessary information related to the quality management system of the manufacturer, including facilities and the reference to the relevant documentation, of the products concerned and the assessments performed, are stated in the Certification Notice which forms an integrative part of this certificate.

This certificate is valid until: 1 October 2023
Issued for the first time: 3 October 2012
Reissued: 10 October 2018

DEKRA Certification B.V.



drs. G.J. Zoetbrood
Managing Director



J.A. van Vugt
Certification Manager

© Integral publication of this certificate and adjoining reports is allowed

DEKRA Certification B.V. is Notified Body with ID no 0344

DEKRA Certification B.V. Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 6802 ED Arnhem, The Netherlands
T +31 88 96 83000 F +31 88 96 83100 www.dekra-certification.com Company registration 09085396

ADDENDUM

Belonging to certificate: 2149850CE01

1/1

CE MARKING OF CONFORMITY MEDICAL DEVICES

Cardiac AF analyzer

Issued to:

Applied Biomedical Systems B.V.
Oxfordlaan 55
6229 EV Maastricht
The Netherlands

This certificate covers the following product(s):

MyDiagnostick, Model 1001R (Class IIa)

Initial date: 3 October 2012

DEKRA Certification B.V.

A blue ink signature of G.J. Zoetbrood, written in a cursive style.

drs. G.J. Zoetbrood
Managing Director

A blue ink signature of J.A. van Vugt, written in a cursive style.

J.A. van Vugt
Certification Manager

© Integral publication of this certificate and adjoining reports is allowed

DEKRA Certification B.V. is Notified Body with ID no 0344

DEKRA Certification B.V. Meander 1051, 6825 MJ Arnhem P.O. Box 5185, 6802 ED Arnhem, The Netherlands
T +31 88 96 83000 F +31 88 96 83100 www.dekra-certification.com Company registration 09085396